

# EU DECLARATION OF CONFORMITY

According to Art. 19 of Regulation (EU) 2017/745 on Medical Devices

**Manufacturer:** Hubei Ingenuity Medical Products Co., Ltd.  
No. 1 West of Industrial Park Road, Yanglinwei Town,  
433000 Xiantao City, Hubei Province, China

**Trademark:** 

**SRN:** Not available yet

**European Representative:** MedPath GmbH  
Mies-van-der-Rohe-Strasse 8  
80807 Munich, Germany

**SRN:** Not available yet

**Trade name:** Medical face mask

**Product Name:** Disposable medical face mask

**Product code / Catalogue number:** INM-3PM/4PM 17.5x9.5cm, 14.5x9.5cm

**Basic UDI:** 697442339

**Classification acc. to MDR Ax. VIII:** Class I, rule I

**Applied Standard & Common Specification:** EN 14683:2019

**Conformity assessment procedure:** Annex II + Annex III of MDR

**Registration number:** DE/CA61/00167698

We, the manufacturer, herewith declare under our sole responsibility that the above-mentioned products meet the provisions of the Regulation (EU) 2017/745 on Medical Devices (MDR), and the provision of the Regulation (EU) 2016/425 on personal protective equipment. All supporting documentations are retained under the premises of the manufacturer.

Signature:

  
Yu Fu

General Manager

Xiantao, 21. 09. 2020





MedPath

## Appendix A: Product Category(ies)

<b>Name</b>
Medical face mask
Surgical gown
Shoe covers
Sleeve covers
Caps
Scrub suit
Bed sheet
Colonoscopy pants

**MedPath GmbH**  
Mies-van-der-Rohe-Strasse 8 · D-80807 München  
Tel.089-189174474 · Fax 089-54858884







MedPath

## EU- Compliance Review Certificate

Regulation (EU) 2017/745 on Medical Devices (MDR), Article 19

Class I non-sterile and without measuring function

No. A0476K001

Manufacturer: Hubei Ingenuity Medical Products Co., Ltd  
No. 1, West of Industrial Park Road,  
Yanglinwei Town, Xiantao City,  
Hubei Province, China

Product See Appendix A

Category(ies):



This is to certify that, MedPath GmbH has verified the aforementioned manufacturer declares to conform with all applicable requirements set in the Medical Device Regulation (EU) 2017/745 for the product category(ies) listed in Appendix A. The manufacturer has provided MedPath GmbH with the appropriate Declaration(s) of Conformity confirming that the medical device(s) fulfills/fulfill the applicable requirements of the aforementioned Regulation. When the manufacturer affixes CE mark to the product category(ies) listed above, it remains the manufacturer's responsibility to ensure continuous compliance with all applicable requirements of the aforementioned Regulation.



**MedPath GmbH**  
Mies-van-der-Rohe-Strasse 8 · D-80807 München  
Tel. 089-189174474 · Fax 089-54858884

Issued Date: 2020-09-12 MedPath GmbH

Expiration Date: 2025-09-12



中国认可  
国际互认  
检测  
TESTING

CNAS 19612

**Test Report**

No.: GOL35EHP062025L1 Issued Date: 2020-10-16

Page 1 of 13

Applicant: HUBEI INGENUITY MEDICAL PRODUCTS CO.,LTD

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Name: Disposable Medical Face Mask

Sample Specification: 175mm×95mm

Manufacturing Date or Lot No.: 2020-09-28/202010

Model: INM-3PM

Manufacturer/Tested company: HUBEI INGENUITY MEDICAL PRODUCTS CO.,LTD

Trade Mark: INMED

Sample Received Date: 2020-09-30

Testing Period: 2020-09-30~2020-10-16

Test Requested: BS EN 14683:2019

Testing Results: Please refer to next page(s)

Edited by:

Checked by:

Approved by:



The page below is blank.



**Test Report**

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 2 of 13

**Summary of Test Results**

No.	Test Item	Test Standard	Performance Requirement Type IIR	Judgement
1	Bacterial Filtration Efficiency Test (BFE),%	BS EN 14683:2019 Appendix B	$\geq 98$	Pass
2	Differential pressure Test (Pa/cm <sup>2</sup> )	BS EN 14683:2019 Appendix C	$< 60$	Pass
3	Splash resistance pressure Test (kPa)	ISO 22609:2004	$\geq 16.0$	Pass
4	Microbial cleanliness Test (CFU/g)	BS EN 14683:2019 Appendix D	$\leq 30$	Pass

Note: Pass = Meet customer requirements;

Fail = Fail customer requirements.

The page below is blank.



**Test Report**  
Test Results

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 3 of 13

5.2.2		TABLE: Bacterial filtration efficiency (BFE)						
Sample No.:	Test Specimen No.:	Dimension of the test specimen L x W (mm x mm)	test area	Flow rate (l/min)	Mean of the total plate counts of the two positive controls	Mean plate count of the negative controls	BFE for each test specimen (%)	Remarks
P062025L1	1	175×95	Φ11cm	28.3	2107	0	99.76	/
	2	175×95	Φ11cm	28.3	2107	0	99.86	/
	3	175×95	Φ11cm	28.3	2107	0	99.72	/
	4	175×95	Φ11cm	28.3	2107	0	99.95	/
	5	175×95	Φ11cm	28.3	2107	0	99.91	/

**Supplementary information:**

- Each specimen was conditioned at 21 °C and 85 % relative humidity for 4 h to bring them into equilibrium with atmosphere prior to testing.
- The side of the test specimen was facing towards the challenge aerosol: face

Test procedure

- Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately  $5 \times 10^5$  CFU/mL.
- Adjust the flow rate through the sampler to 28.3 L/min.
- Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- Perform a positive control run without a test specimen to determine the number of viable aerosol particles



**Test Report**

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 4 of 13

being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.

1.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer.

Immediately begin sampling the aerosol using the sampler.

1.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.

1.4.3 Time the air pressure and sampler to run for 2 min.

1.4.4 At the conclusion of the positive control run, remove plates from the sampler.

1.5 Place new agar plates into sampler and clamp the test specimen into the top of the sampler, with the inside of the specimen facing towards the bacterial challenge (test area:  $\varnothing 11\text{cm}$ ).

1.6 Repeat the challenge procedure for each test specimen.

1.7 Repeat a positive control after completion of the sample set.

1.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.

1.9 Incubate agar plates at  $(37\pm 2)^{\circ}\text{C}$  for  $(48\pm 4)$  h.

1.10 Count each of the six-stage plates of the sampler.

The page below is blank.



**Test Report**  
Test Results

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 5 of 13

5.2.3	TABLE: Breathability (Differential pressure)				
Sample No.:	Test Specimen number-Test area number	Differential pressure for each test area (Pa/cm <sup>2</sup> )	The averaged differential pressure for each test specimen (Pa/cm <sup>2</sup> )	Flow rate (l/min)	Remarks
P062025L1	1-1	27.3	24.1	8	/
	1-2	22.1		8	/
	1-3	22.7		8	/
	1-4	23.9		8	/
	1-5	24.7		8	/
	2-1	30.3	26.2	8	/
	2-2	25.0		8	/
	2-3	29.5		8	/
	2-4	19.4		8	/
	2-5	26.9		8	/
	3-1	23.3	22.2	8	/
	3-2	22.1		8	/
	3-3	20.7		8	/
	3-4	22.9		8	/
	3-5	22.1		8	/



**Test Report**  
Test Results

No.: GOL35EHP062025L1


Issued Date: 2020-10-16 Page 6 of 13

5.2.3 TABLE: Breathability (Differential pressure)					
Sample No.:	Test Specimen number-Test area number	Differential pressure for each test area (Pa/cm <sup>2</sup> )	The averaged differential pressure for each test specimen (Pa/cm <sup>2</sup> )	Flow rate (l/min)	Remarks
P062025L1	4-1	28.9	26.1	8	/
	4-2	26.8		8	/
	4-3	27.0		8	/
	4-4	24.2		8	/
	4-5	23.8		8	/
	5-1	30.2	25.0	8	/
	5-2	26.5		8	/
	5-3	23.2		8	/
	5-4	21.5		8	/
	5-5	23.6		8	/

**Supplementary information:**

Each specimen was conditioned at 21 °C and 85 % relative humidity for 4 h to bring them into equilibrium with atmosphere prior to testing.

Sketch map of test location:





**Test Report**

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 7 of 13

Test procedure

1.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.

1.2 The pretreated specimen is placed across the orifice (total area 4.9cm<sup>2</sup>, test area diameter 25mm) and clamped into place so as to minimize air leaks.

1.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.

1.4 The differential pressure is read directly.

1.5 The procedure described in steps 1.1-1.4 is carried out on 5 different areas of the mask and calculate averaged.

The page below is blank.



**Test Report**  
Test Results

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 8 of 13

5.2.4	TABLE: Splash resistance pressure			
Sample No.:	Test mask No.:	The material of tested mask	Test result (Pass/fail)	Remarks
P062025L1	1	/	Pass	/
	2	/	Pass	/
	3	/	Pass	/
	4	/	Pass	/
	5	/	Pass	/
	6	/	Pass	/
	7	/	Pass	/
	8	/	Pass	/
	9	/	Pass	/
	10	/	Pass	/
	11	/	Pass	/
	12	/	Pass	/
	13	/	Pass	/
	14	/	Pass	/
	15	/	Pass	/
	16	/	Pass	/



**Test Report**  
Test Results

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 9 of 13

5.2.4	TABLE: Splash resistance pressure			
Sample No.:	Test mask No.:	The material of tested mask	Test result (Pass/fail)	Remarks
P062025L1	17	/	Pass	/
	18	/	Pass	/
	19	/	Pass	/
	20	/	Pass	/
	21	/	Pass	/
	22	/	Pass	/
	23	/	Pass	/
	24	/	Pass	/
	25	/	Pass	/
	26	/	Pass	/
	27	/	Pass	/
	28	/	Pass	/
	29	/	Pass	/
	30	/	Pass	/
31	/	Pass	/	
32	/	Pass	/	

**Supplementary information:**

- Each specimen was conditioned at 21 °C and 85 % relative humidity for 4 h to bring them into equilibrium with atmosphere prior to testing.
- The description of target area tested: The center of outside
- Any technique used to enhance visual detection of synthetic blood: cotton swab
- The temperature and relative humidity for testing: 21 °C and 85 %
- Description of any pre-treatment techniques used: constant temperature and humidity machine was used



## Test Report

No.: GOL35EHP062025L1

Issued Date: 2020-10-16

Page 10 of 13

### Test procedure

- 1.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 1.2 Determine the density of the synthetic blood.
- 1.3 Fill the reservoir with new synthetic blood.
- 1.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 1.5 Set the reservoir pressure to the approximate pressure.
- 1.6 Place the targeting plate approximately 1 cm away from the mask.
- 1.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 1.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 1.9 Calculate the difference in weight of the two spurts.
- 1.10 Adjust the reservoir pressure and repeat steps 1.7 to 1.9 until the weight difference is within the target range.
- 1.11 Record the weight difference for the spurts exiting the nozzle.
- 1.12 Record the pressure in the reservoir.
- 1.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 1.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 1.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2%~-5% of the difference in weight from the nozzle.



**Test Report**

No.: GOL35EHP062025L1

Issued Date: 2020-10-16 Page 11 of 13

1.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.

1.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 1.7 to 1.11).

1.18 Adjust the timer setting until 2 ml of fluid passes through the hole for three spurts in a row. For a test fluid with a density of 1,005 g/cm<sup>3</sup>, the output should weigh 2,01 g.

1.19 Record the timer setting to use as the starting point for subsequent testing.

1.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.

1.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.

1.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.

1.23 Report the results (none/penetration) for each test specimen at the test pressure.

The page below is blank.



**Test Report**  
Test Results

No.: GOL35EHP062025L1

Issued Date: 2020-10-16

Page 12 of 13

5.2.5	TABLE: Microbial cleanliness (Bioburden)			
Sample No.:	Mask(under test) No.:	Weight of each mask (g)	Total bioburden per individual mask (CFU/g)	Remarks
P062025L1	1	3.4	<1	/
	2	3.4	<1	/
	3	3.3	<1	/
	4	3.4	<1	/
	5	3.3	<1	/
Supplementary information: /				

Test procedure

1.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.

1.2 The mask is weighed and aseptically removed from the packaging and placed in a aseptic container containing 300 mL of extraction liquid, then mix well.



1.3 After extracting, 100 mL of the extraction liquid is filtered through a 0.45 µm filter, wash with 100mL sterile 0.1% peptone water (containing 2g/L Tween 20) each time. Then transfer the filter aseptically to the TSA plate to count the active aerobic microorganisms. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungus count.

1.4 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.

1.5 Calculate the colonies of each agar plate.



Photo:



Pony authenticate the photo on original report only

——End of Report——





## 声明 Statement

1. 本报告无检验检测专用章、报告骑缝章和批准人签章无效。  
This report is invalid without special seal of inspection, cross-page seal and the approver's signatures.
2. 本报告页面所使用“PONY”、“谱尼”字样为本单位的注册商标, 其受《中华人民共和国商标法》保护, 任何未经本单位授权的擅自使用和仿冒、伪造、变造“PONY”、“谱尼”商标均为违法侵权行为, 本单位将依法追究其法律责任。  
The pattern and characters of "PONY" and "谱尼" used in this report are protected by the trademark law of the People's Republic of China. Any unauthorized usage, counterfeit, forgery and alteration of trademarks of "PONY" and "谱尼" are the violations of the law. The PONY has the right to pursue all legal liabilities of the subject of the delict.
3. 委托单位对报告数据如有异议, 请于报告完成之日起十五日内(初级农产品报告请于报告收到之日起五日内)向本单位书面提出复测申请, 同时附上报告原件并预付复测费。  
If the applicant has any questions about the results, shall provide a written retest application with the original report, and prepay the retest fees to PONY within fifteen days since the approval date (as an exception, it shall be within five days since the date received for the primary agriculture products report).
4. 委托单位办理完毕以上手续后, 本单位会尽快安排复测。如果复测结果与异议内容相符, 本单位将退还委托单位的复测费。  
After the applicant finishes the procedure mentioned above, PONY shall arrange the retest as soon as possible. If the retest result accords with the applicant dissent, PONY shall refund the retest fees.
5. 不可重复性或不能进行复测的实验, 不进行复测, 委托单位放弃异议权利。  
Tests that can not be repeated and tested shall not be carried out again.
6. 委托单位对样品的代表性和资料的真实性负责, 否则本单位不承担任何相关责任。  
The applicant should undertake the responsibility for the provided samples' representativeness and document authenticity. Otherwise, PONY has not any relevant responsibilities.
7. 本报告仅对所测样品负责, 报告数据仅反映对所测样品的评价, 对于报告及所载内容的使用、使用所产生的直接或间接损失及一切法律后果, 本单位不承担任何经济和法律后果。  
This report is only responsible for the provided sample. The test results only represent the evaluation of the tested sample. PONY will not be responsible for any economical or legal liability generated from direct or indirect usage of the test report.
8. 本单位有权在完成报告后按规定方式处理所测样品。  
PONY has the right to dispose the tested sample by rules, after approval of the test report.
9. 本单位保证工作的客观公正性, 对委托单位的商业信息、技术文件等商业秘密履行保密义务。  
PONY assures objectivity and impartiality of the test, and fulfills the obligation of confidentiality for applicant's commercial information, and technique document.
10. 本报告私自转让、盗用、冒用、涂改、未经本单位批准的复制(全文复制除外)或以其它任何形式的篡改均属无效, 本单位将对上述行为追究其相应的法律责任。  
The report is invalid in case of illegal transfer, embezzlement, imposture, modification or any altering, reproducing except in full, without approval of PONY. PONY shall investigate and affix the applicant's legal liability accordingly.

### ▲ 防伪说明 (Anti-counterfeiting Description):

- (1) 报告编号是唯一的;  
The test report has exclusive report code.
- (2) 报告采用特制防伪纸张印制, 纸张表面带有"PONY"防伪纹路, 该防伪纹路不支持复印, 即复制件不会带有"PONY"防伪纹路。  
The test report is printed by anti-copying paper whose surface shows "PONY" security print with specific anticounterfeiting technique. Security print will disappear after copying. Duplicates are not expected to give "PONY" security print under any circumstances.



全国服务热线  
400-819-5688

WWW.PONYTEST.COM

扫描二维码  
关注谱尼测试微信  
公众号 PONY4008195688



北京实验室: (010) 83055000	武汉实验室: (027) 83997127	哈尔滨实验室: (0451)58627755
上海实验室: (021) 64851999	长春实验室: (0431)85150908	石家庄实验室: (0311)85376660
青岛实验室: (0532)88706866	大连实验室: (0411)87336618	乌鲁木齐实验室: (0991) 6684186
深圳实验室: (0755)26050909	郑州实验室: (0371)69350670	呼和浩特实验室: (0471) 3450025
天津实验室: (022) 23607888	西安实验室: (029) 89608785	杭州实验室: (0571)85806807
苏州实验室: (0512)62997900	太原实验室: (0351) 7555762	宁波实验室: (0574)87977185
		温州实验室: (0577)88271060
		合肥实验室: (0551)63843474
		广州实验室: (020) 89224310
		厦门实验室: (0592)5568048
		成都实验室: (028) 87702708